



आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India  
**QUOTATION/ RATE INQUIRY**

1. Quotation is hereby invited on behalf of Director, ICMR- National Institute of Malaria Research from **M/S. Solution One** authorized dealer of **M/S. Thermofisher Scientific** for supply of **Multiskan Sky**.

**Specifications: -**

1. Spectral scanning unit with for UV-visible-NearIR wavelength range with dedicated cuvette port, 96 and 384 micro-well plate format
2. Capable to read at low volume samples of 2µl-10µl using low volume analysis plate in directly with standalone mode
3. Wavelength range from 200nm to 1000nm with 1 nm steps
4. Spectral scanning speed: 6 secs for 96-well microplate 10 secs for 384-well microplate
5. Performance Specifications: Bandwidth: < 2.5 nm or better and Xenon flash lamp life should be for 10 million 96 well microplates
6. The instrument should have inbuilt incubation and linear shaking options for longer kinetic assays
7. Incubation temperature: from ambient + 2°C to 45°C
8. Should have area selection option, for different measuring parameters for different area in a same plate
9. Visualize data in both numerical mode and heat-map/virtual image of plate
10. The instrument should have minimum 2 USB ports, one for the easy data transfer and able to connect Wi-Fi dongle
11. Capable to work as a standalone system without computer and also able to run with computer controlled software
12. Capable to read end point, kinetics, spectral scanning and also kinetic spectral scan
13. Analysis Software supplied should be supplied with unlimited user license
14. System should have Power Save function for reduced energy consumption when the instrument is 'on' but not in use

**2. The Notice is issued for general information of Aspirant Manufacturer/ Dealer/ Distributor to submit their objections/ proposal, if any, on proprietorship of these items in terms of Rule 166 (i) of GFR, 2017.**

3. In case, the product of any other Manufacturer/ Authorized distributor/ dealer conforms to the aforesaid items, they may submit their proposal for the supply of same item along with the following: - (a) Item brochure (b) Point-by-point compliance, along with all relevant documentary evidence.

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**icmr** | **NIMR**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH | NATIONAL INSTITUTE OF  
MALARIA RESEARCH

आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

4. The objection/ proposal should be sent through email to "[quotations.nimr@gmail.com](mailto:quotations.nimr@gmail.com)" so as to reach on or before **04.02.2021** up to **03:00 PM**, failing which it will be presumed that no any other vendor is interested to offer comments/ objection and case will be decided accordingly on its merit.

Sd/-x-x-x  
Store-in-charge  
for Director

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

### Terms & Conditions:

1. Quotation send by email will only be considered.
2. The quotation is to be submitted only in the format enclosed. Not acceptable in any other form.
3. The bidder should clearly mention whether they are Manufacturer or authorized Dealer/ Agent/ Distributers/ Suppliers of the Manufacturer. In the case of Dealer/ Agent/ Distributers/Suppliers, valid authorization letter from the Manufacturer address to Director, ICMR-NIMR, New Delhi should be submitted along with the quotation. One manufacturer can authorize only one agent/ dealer.
4. Quotation received after the last date & time is liable to be rejected.
5. GST will be applicable as per GoI norms to ICMR-NIMR, New Delhi.
6. The bidder conditions mentioned in the quotation shall not be binding on ICMR-NIMR, New Delhi.
7. Quotation should be preferably, typewritten and every correction in the quotation should invariably be signed by the bidder, failing which, the quotation is liable to be rejected.
8. Any dispute concerning any Terms and Conditions of the Quotation and/ or the supply of items, will be subject to Delhi Jurisdiction only.
9. Price should be quoted on the Vendor's Letter Head, duly stamped and signed.
10. Materials should be delivered within 4 weeks at ICMR-NIMR, New Delhi from the date of issue of purchase order. The liquidated charges @ 1% per week shall be imposed if supply made after expiry of delivery period subject to maximum 10% of the total value of goods/ contract value.
11. In the quotation, S No. and name of the items should be same as per the format enclosed.
12. Validity of quotation should be for a minimum period of 6 months.
13. If, in the price structure quoted by a bidder, there is discrepancy between the unit price and the total price (which is obtained by multiplying the unit price by the quantity), the unit price shall prevail and the total price corrected accordingly. If there is a discrepancy between the amount expressed in words and figures, the amount in words shall prevail.
14. Director, NIMR reserves the right to reject any or all quotations without assigning any reason.
15. Please clearly mention the quotation No. & due date in the subject line of email.
16. Price of quotations must be included with 3 years Comprehensive & 2 years Non-Comprehensive Warranty.

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**icmr** | **NIMR**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH | NATIONAL INSTITUTE OF  
MALARIA RESEARCH

आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

17. Performance Security: - The vendor shall furnish a performance security in form of an Accounts payee bank draft/ pay order, Fixed Deposit Receipt (FDR) from scheduled commercial bank OR Bank Guarantee (PBG) from a scheduled commercial bank in India in an acceptable form safeguarding the purchase interest in all respects. The Bank guarantee should cover the warranty period as mentioned above from the date of installation of the equipment. The value of the Bank guarantee is 10% of the order value.

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

### FINANCIAL BID FORMAT (On Letter Head)

To, Director,  
ICMR-NIMR,  
New Delhi

Offer for supply of Multiskan Sky against NIMR inquiry No.2/59/2020-21/Store is under:

S No	Particulars	Req Qty	Company/ Brand Name	Unit Price	Discount in %	GST @ 5%	Price to NIMR (Including GST & Discount)	Net Price (Qty * Price to NIMR)
1.	Multiskan Sky	1						

I/ We, the undersigned, hereby declare that:

1. I/We certify that quoted product meets the specification as described in the call for quotation.
2. I/ We certify that rate quoted for this product does not exceed the Maximum Retail Price of the product.

Name and Surname  
Signature & Stamp of Bidder

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

## दर सूची/ दर संबंधी पृछताछ

1. निदेशक, आईसीएमआर-राष्ट्रीय मलेरिया अनुसंधान संस्थान की ओर से एतद् द्वारा मैसर्स **M/S. Solution One** मैसर्स के अधिकृत डीलर **M/S. Thermofisher Scientific** से **Multiskan Sky** की आपूर्ति के लिए दर सूची आमंत्रित की जाती है।

### **Specifications: -**

1. Spectral scanning unit with for UV-visible-NearIR wavelength range with dedicated cuvette port, 96 and 384 micro-well plate format
2. Capable to read at low volume samples of 2µl-10µl using low volume analysis plate in directly with standalone mode
3. Wavelength range from 200nm to 1000nm with 1 nm steps
4. Spectral scanning speed: 6 secs for 96-well microplate 10 secs for 384-well microplate
5. Performance Specifications: Bandwidth: < 2.5 nm or better and Xenon flash lamp life should be for 10 million 96 well microplates
6. The instrument should have inbuilt incubation and linear shaking options for longer kinetic assays
7. Incubation temperature: from ambient + 2°C to 45°C
8. Should have area selection option, for different measuring parameters for different area in a same plate
9. Visualize data in both numerical mode and heat-map/virtual image of plate
10. The instrument should have minimum 2 USB ports, one for the easy data transfer and able to connect Wi-Fi dongle
11. Capable to work as a standalone system without computer and also able to run with computer controlled software
12. Capable to read end point, kinetics, spectral scanning and also kinetic spectral scan
13. Analysis Software supplied should be supplied with unlimited user license
14. System should have Power Save function for reduced energy consumption when the instrument is 'on' but not in use

2. जीएफआर, 2017 के नियम 166 (i) के संदर्भ में इन वस्तुओं के स्वामित्व पर, अपनी आपत्तियां / प्रस्ताव प्रस्तुत करने के लिए इच्छुक निर्माता / डीलर / वितरक की सामान्य जानकारी के लिए नोटिस जारी किया जाता है।
3. यदि किसी अन्य निर्माता/प्राधिकृत वितरक/डीलर का उत्पाद पूर्वकथित मर्दों के अनुरूप है तो वे निम्नलिखित के साथ एक ही वस्तु की आपूर्ति के लिए अपना प्रस्ताव प्रस्तुत कर सकते हैं: - (क) मद विवरणिका (ख) सभी संगत दस्तावेजी साक्ष्यों के साथ बिंदुवार अनुपालन।

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**icmr** | **NIMR**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH | NATIONAL INSTITUTE OF  
MALARIA RESEARCH

आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

4. आपत्ति/ प्रस्ताव "[quotations.nimr@gmail.com](mailto:quotations.nimr@gmail.com)" ईमेल पर भेजा जाना चाहिए ताकि दिनांक **04.02.2021** को या उससे पूर्व **03:00 PM** पहुंच सके अन्यथा यह माना जाएगा कि कोई भी अन्य विक्रेता टिप्पणियां/आपत्ति प्रस्तुत करने के इच्छुक नहीं है तथा मामला तदनुसार योग्यता क्रम के आधार पर तय किया जाएगा।

एसडी/ख-ख-ख  
भंडार प्रभारी  
कृते निदेशक

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**icmr** | **NIMR**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH | NATIONAL INSTITUTE OF  
MALARIA RESEARCH

आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research  
Ministry of Health and Family Welfare, Government of India

## नियम एवं शर्तें:

1. केवल ईमेल द्वारा भेजी गई दर सूची पर ही विचार किया जाएगा।
2. दर सूची केवल संलग्न प्रारूप में ही प्रस्तुत करनी हैं। किसी अन्य रूप में स्वीकार्य नहीं हैं।
3. बोलीदाता को स्पष्ट रूप से उल्लेख करना चाहिए कि वे निर्माता हैं या प्राधिकृत डीलर/एजेंट/वितरक/निर्माता के आपूर्तिकर्ता हैं। डीलर/एजेंट/वितरकों/आपूर्तिकर्ताओं के मामले में, निर्माता के पते से निदेशक, आईसीएमआर-एनआईएमआर, नई दिल्ली को मान्य प्राधिकरण पत्र को दर सूची के साथ प्रस्तुत किया जाना चाहिए। एक निर्माता केवल एक एजेंट/डीलर को ही प्राधिकृत कर सकता है।
4. अंतिम तिथि और निर्धारित समय के बाद प्राप्त दर सूची अस्वीकार की जा सकती है।
5. आईसीएमआर-एनआईएमआर, नई दिल्ली के लिए जीएसटी भारत सरकार के नियमानुसार लागू होगा।
6. दर सूची में बोलीदाता की उल्लेखित शर्तें आईसीएमआर-एनआईएमआर, नई दिल्ली के लिए बाध्यकारी नहीं होंगी।
7. दर सूची विशेषतः, टंकित होनी चाहिए और दर सूची में प्रत्येक सुधार को अनिवार्य रूप से बोलीदाता द्वारा हस्ताक्षरित होना चाहिए अन्यथा दर सूची को अस्वीकार कर दिया जाएगा।
8. दर सूची और/या वस्तुओं की आपूर्ति के किसी भी नियम और शर्तों से संबंधित कोई भी विवाद केवल दिल्ली क्षेत्राधिकार के अधीन होगा।
9. मूल्य वस्तुतः विक्रेता के पत्र शीर्ष पर उद्धृत विधिवत मोहर के साथ हस्ताक्षरित होना चाहिए।
10. सामग्री ऑर्डर जारी करने की तारीख से आईसीएमआर-एनआईएमआर, नई दिल्ली में-----सप्ताह के भीतर पहुंच जाना चाहिए। सुपुर्द करने की अवधि की समाप्ति के बाद की गई आपूर्ति पर प्रति सप्ताह 1% का निर्णित प्रभार लगाया जाएगा जोकि सामान/संविदा मूल्य के कुल मूल्य का अधिकतम 10% के अधीन है।
11. दर सूची में क्रम संख्या एवं मदों के नाम संलग्न प्रारूप के अनुसार होने चाहिए।
12. दर सूची की वैधता न्यूनतम 6 महीने होनी चाहिए।
13. यदि किसी बोलीदाता द्वारा उद्धृत मूल्य संरचना में, इकाई मूल्य और कुल मूल्य (जो मात्रा को इकाई मूल्य से गुणा करके प्राप्त किया जाता है) के बीच विसंगति है, तो इकाई मूल्य पर विचार किया जाएगा

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**icmr** | **NIMR**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH | NATIONAL INSTITUTE OF  
MALARIA RESEARCH

आई.सी.एम.आर.-राष्ट्रीय मलेरिया अनुसंधान संस्थान  
स्वास्थ्य अनुसंधान विभाग,  
स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार  
ICMR-National Institute of Malaria Research  
Department of Health Research

Ministry of Health and Family Welfare, Government of India

- और तदनुसार कुल मूल्य सही होगा। यदि शब्दों और अंकों में व्यक्त राशि के बीच कोई विसंगति है तो शब्दों में राशि पर विचार किया जाएगा। निदेशक, एनआईएमआर बिना कोई कारण बताए किसी भी या सभी दर सूची को अस्वीकार करने का अधिकार सुरक्षित रखता है।
14. निदेशक, एनआईएमआर के पास बिना कोई कारण बताए किसी भी या सभी संविदाओं को अस्वीकार करने का अधिकार सुरक्षित है।
  15. कृपया ईमेल के विषय के अंतर्गत दर सूची की संख्या और नियत तारीख का स्पष्ट उल्लेख करें।
  16. दर सूची मूल्य के अंतर्गत 3 साल की व्यापक और 2 साल की गैर-व्यापक वारंटी के साथ शामिल होना चाहिए।
  17. निष्पादन प्रतिभूति:- विक्रेता द्वारा भारत में सभी प्रकार के खरीद रुझान/रुचि की सुरक्षा के लिए एक स्वीकार्य प्रपत्र में किसी अनुसूचित वाणिज्यिक बैंक से बैंक गारंटी या अनुसूचित वाणिज्यिक बैंक से सावधि जमा रसीद/एक अकाउंट पेयी बैंक ड्राफ्ट/पे ऑर्डर के रूप में एक निष्पादन प्रतिभूति प्रस्तुत करनी होगी। उपस्कर की संस्थापन दिनांक से उल्लेखित बैंक गारंटी में वारंटी अवधि को शामिल किया जाना चाहिए। बैंक गारंटी का मूल्य, आदेश मूल्य का 10% है।

हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।



**वित्तीय बोली प्रारूप (पत्र शीर्ष में)**

सेवा में

निदेशक

आईसीएमआर-एनआईएमआर,  
नई दिल्ली

Multiskan Sky की आपूर्ति के लिए प्रस्ताव एनआईएमआर पृछताछ संख्या 2/59/2020-21/Store के संदर्भ में ब्यौरा नीचे दिया गया है:

क्रम सं.	विवरण	अपेक्षित मात्रा	कंपनी/ब्रांड का नाम	इकाई मूल्य	छूट (% में)	जीएसटी @ 5%	एनआईएमआर के लिए मूल्य (जीएसटी एवं छूट सहित)	कुल मूल्य(गुणवत्ता* मूल्य एनआईएमआर)
1.	Multiskan Sky	1						

मैं/हम, अद्योहस्ताक्षरी एतद् द्वारा घोषणा करते हैं कि:

1. मैं/हम प्रमाणित करते हैं कि दर सूची में दर्शाए विनिर्देशों अनुसार ही उद्धृत उत्पाद को पूरा करते हैं।
2. मैं/हम प्रमाणित करते हैं कि इस उत्पाद की उद्धृत दर उत्पाद की अधिकतम खुदरा मूल्य से अधिक नहीं है।

नाम और उपनाम  
हस्ताक्षर एवं बोलीदाता की मोहर


हमारे संस्थान में आपके हिन्दी पत्रों का स्वागत है।

## To whom it may concern

This is to certify that Thermo Scientific™ Multiskan Sky™, quoted by us is an Article of proprietary nature, and we are the original manufacturer of the said item. This Multiskan Sky item is solely manufactured by us with following features and not by anyone else in the entire country/ world.

- **Multiskan Sky can be used as stand-alone system without computer and also able to run with computer controlled SkanIt™ software.**
- Instrument is able to read 16 low volume samples of 2µl-10µl using low volume analysis plate for DNA/RNA purity & concentration or protein concentration with stand-alone mode.
- Stand-alone mode using 7-inch touch screen for quick usage.
- Data storage of more than 400 000 end point measurements with single wavelength (measurement of 400 000 pc. of 96-well or 384-well plates).
- Four USB ports to PC and other data transfer devices and one Ethernet port for cloud connection
- Data access from anywhere via cloud capabilities, wired or wireless network connection. Able to connect Wi-Fi dongle
- Instrument is able to provide the wavelength range from **200 nm to 1000 nm** with 1 nm steps.
- Spectral scanning speed: 10 s from 200 to 1000 nm with 1 nm steps per sample.
- Photometric **Bandwidth: < 2.5 nm**
- **Xenon flash lamp life time is typically 10 million 96 well microplates.**
- The instrument has inbuilt **incubation** and **shaking** options for longer time kinetic assays.
- Incubation **temperature: from ambient +2 °C to +45 °C.**
- Measurement speed: 6 s for 96 well plate and 10 s for 384 well plate
- System has Power Save function for reduced energy consumption when the instrument is 'on' but not in use.
- Visualize data in both numerical mode and heat-map/virtual image of plate.
- **Multiple language versions: English, German, French, Italian, Spanish, Portuguese, Russian, Chinese and Japanese**
- System is supplied with unlimited user license SkanIt software for the instrument control and data analysis.
- Single software program allow any number of measurement steps (different detection modes) within the program.
- No loss of already measured data even in case of power failure. User need not wait till end of the protocol to save the data file.
- Software has option for area selection, i.e. different protocols at different area of the same plate for applications such as differential reading or reading in different intervals within same plate.
- This product is certified as per CE/ IEC guideline.

For Thermo Fisher Scientific Oy



Arja Lamberg  
Sr. Product Manager  
Microplate Readers and Washers  
Imaging, Labeling and Detection

